



CE Declaration of Conformity

We: Labconco Corporation
8811 Prospect Avenue
Kansas City, MO 64132 USA
Ph. 816-333-8811 Fax. 816-363-0130
labconco@labconco.com

Authorized Representative
Mauro Ferri c/o BIPIU'BI S.r.l.
Via Vittorio Veneto, 12
20091 Bresso (MI)
Italia

being the representative party solely responsible for this certification, hereby declare that the following products:

Equipment: Laboratory Equipment – PuriCare® Lab Animal Research Stations

Type, Model: 3820020 Open Access Station, 5FT, 230V, 50/60 Hz (US)
3820021 Open Access Station, 4FT, 230V, 50/60 Hz (US)
3820023 Open Access Station, 4FT, 230V, 50/60 Hz (Schuko)
3820024 Open Access Station, 4FT, 230V, 50/60 Hz (UK)
3820026 Open Access Station, 4FT, 230V, 50/60 Hz (China/Australia)
3820028 Open Access Station, 4FT, 230V, 50/60 Hz (India)
3820033 Open Access Station, 5FT, 230V, 50/60 Hz (Schuko)
3820034 Open Access Station, 5FT, 230V, 50/60 Hz (UK)
3820036 Open Access Station, 5FT, 230V, 50/60 Hz (China/Australia)
3820038 Open Access Station, 5FT, 230V, 50/60 Hz (India)

NOTE: Above part numbers may be followed by a 5-digit number for custom modifications.

Manufacturer: Labconco Corporation
8811 Prospect Avenue
Kansas City, MO 64132 USA
Ph. 816-333-8811 Fax. 816-363-0130
labconco@labconco.com

are in accordance with the following directives:

2014/30/EU	Electromagnetic Compatibility Directive	2015/863/EU	RoHS3 Directive
2014/35/EU	Low Voltage Directive		
2006/42/EU	Machinery Directive		
2011/65/EU	RoHS Directive		

are in conformity with the standards listed below:

EN61010-1:2010+A1:2019+A2:2021	EN61000-4-3:2020
EN61326-1:2021	EN61000-4-4:2012+A1:2022
EN55011:2016+A1:2017+A2:2021	EN61000-4-5:2014+A1:2017
EN61000-3-2:2019+A1:2021	EN61000-4-6:2022
EN61000-3-3:2013+A2:2021	EN61000-4-11:2020
EN61000-4-2:2009	EN63000:2018

when installed and operated in accordance with the manufacturers installation and operating instructions.

Brandon Gray
Vice President of Engineering
Labconco Corporation
Kansas City, MO USA

15 OCT 2024
Date of issue



UKCA Declaration of Conformity

We: Labconco Corporation
8811 Prospect Avenue
Kansas City, MO 64132 USA
Ph. 816-333-8811 Fax. 816-363-0130
labconco@labconco.com

being the representative party, in accordance with UK government guidance, solely responsible for this certification, hereby declare that the following product:

Equipment: Laboratory Equipment – PuriCare® Lab Animal Research Stations

Type, Model: 3820020 Open Access Station, 5FT, 230V, 50/60 Hz (US)
3820021 Open Access Station, 4FT, 230V, 50/60 Hz (US)
3820023 Open Access Station, 4FT, 230V, 50/60 Hz (Schuko)
3820024 Open Access Station, 4FT, 230V, 50/60 Hz (UK)
3820026 Open Access Station, 4FT, 230V, 50/60 Hz (China/Australia)
3820028 Open Access Station, 4FT, 230V, 50/60 Hz (India)
3820033 Open Access Station, 5FT, 230V, 50/60 Hz (Schuko)
3820034 Open Access Station, 5FT, 230V, 50/60 Hz (UK)
3820036 Open Access Station, 5FT, 230V, 50/60 Hz (China/Australia)
3820038 Open Access Station, 5FT, 230V, 50/60 Hz (India)

NOTE: Above part numbers may be followed by a 5-digit number for custom modifications.

Manufacturer: Labconco Corporation
8811 Prospect Avenue
Kansas City, MO 64132 USA
Ph. 816-333-8811 Fax. 816-363-0130
labconco@labconco.com

the object of the declaration described above is in conformity with the relevant UK Statutory Instruments (and their amendments):

S.I. 2016/1091 Electromagnetic Compatibility Directive
S.I. 2016/1101 Low Voltage Directive
S.I. 2008/1597 Machinery Directive
S.I. 2012/3032 RoHS Directive

are in conformity with the standards listed below:

EN61010-1:2010+A1:2019+A2:2021	EN61000-4-3:2020
EN61326-1:2021	EN61000-4-4:2012+A1:2022
EN55011:2016+A1:2017+A2:2021	EN61000-4-5:2014+A1:2017
EN61000-3-2:2019+A1:2021	EN61000-4-6:2022
EN61000-3-3:2013+A2:2021	EN61000-4-11:2020
EN61000-4-2:2009	EN63000:2018

when installed and operated in accordance with the manufacturers installation and operating instructions.

Brandon Gray
Vice President of Engineering
Labconco Corporation
Kansas City, MO USA

15 OCT 2024
Date of issue